



MINISTRY OF HEALTH

Telephone: Nairobi 254-020-2717077
Fax: 254-2719008
Email: ps.medical@health.go.ke

AFYA HOUSE
CATHEDRAL ROAD
P. O Box 30016-00100
NAIROBI

When replying please quote:

Ref No.: MOH/SDMS/EACE/OCBI/001/2025-2026

Date: 10th July, 2026

CLARIFICATION OF TENDER DOCUMENTS: ADDENDUM NO. 1 – 10th JULY 2026

SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING, TESTING AND ON USE TRAINING OF BIOMEDICAL EQUIPMENT

TENDER NO. MOH/SDMS/EACE/OCBI/001/2025-2026

Reference is made to the above tender that appeared on the Ministry's website on 12th June, 2026. Interested bidders have sought for some clarifications on the tender document which are hereby made as follows:


S/NO	DESCRIPTION OF INQUIRY	CLARIFICATION
1	I write to seek clarification on Item 21 Lot 10 (LOT 10-21) Mobile X-Ray Unit. Whereas all major equipment request by the FDA or CE, this particular equipment-high value equipment, has been designated on clause number 16 as it should have quality certification CE and FDA. Please clarify if it should be FDA or CE instead of FDA and CE.	The required quality certifications is FDA or CE.
2	Kindly confirm that this is an International Tender and that international bidders are permitted to submit their bids in USD.	The procurement procedure is Open Competitive bidding (International) (OCBI) Check section II Bid Data Sheet (ITB 1.1) The bidder may express the Bid prices in any currency if they wish to be paid in a combination of amounts in different currencies but shall use NO more than three currencies in addition to the currency of purchaser country as provided under paragraph 15.2 of ITB.

3	Kindly confirm that the bid evaluation will be conducted on a lot-wise basis and that each selected lot will be awarded to a single bidder.	The tender will be evaluated on a Lot-by-Lot basis and only a single bidder can be awarded any of the Lots
4	Kindly provide the complete list of installation sites , along with the quantity of each equipment item to be delivered and installed at each respective site	The project site(s) Destination(s) is East Africa Centers of Excellence (EACE) Complex at Kenyatta National Hospital Grounds off Ngong Road.
5	Kindly confirm whether CE or FDA approval will be considered acceptable as proof of product compliance.	The required quality certifications is FDA or CE .
6	With reference to Lot 12 – Waste Management Unit , kindly confirm the required quantity.	Waste Management unit should be one unit.
7	With reference to Lot 8 – Operation Theatres , Item: Anesthetic Machines , kindly clarify the full form of " PNS ".	Peripheral nerve stimulator. Bidders are therefore advised to ignore all the specifications related to that item as the same will NOT be evaluated.
8	We are preparing a bid for Lot 7 (Renal Hemodialysis) and need clarification on the specification for Lot 7-2, Pediatric Dialysis Machines. The specification asks for a machine that treats patients weighing up to 10 kg and also has on-line HDF capability. We have checked with the manufacturers we represent and are not aware of any HDF machine on the market indicated for patients below 10 kg. Was the intention a minimum patient weight of 10 kg? Please confirm. The specification also calls for a double blood pump. Please confirm if this means a dual-pump setup for single-needle dialysis.	The intended minimum patient weight is 10kgs. The requirement for on line HDF capacity applies to pediatric patients from 10kgs and above, in line with current international practice and manufacturer indications. "Double blood pump' requirement A dual pump configuration designed for single needle dialysis.
9	Joint venture consortium You have stated on clause ITB 4.1 (c) Maximum number of members in the Joint Venture, Consortium or Association (JV) shall be: Three (3). Clause ITB 4.1 (d) Minimum	

	share of a member of Joint Venture, Consortium or Association (JV) in the contract shall not be less than 25% of the total value of the contract. Take an example of Lot 10, which has 25 medical equipment, with clause ITB 4.1 where each member has a minimum of 25%, requires each member (3 maximum), to bid a minimum of 6 equipment. We would like therefore to clarify: 1. If we have a member who can offer 1 (one) specific critical equipment within the lot, required to fulfil the lot, can the Joint Venture consortium be considered valid? 2. We propose that you allow this minimum share to be considered to allow even for a member with one equipment to join a consortium, as opposed to the minimum of 25% of the Lot.	No .Refer to ITB 14.6 No. Minimum share of a member of joint venture, consortium or association (JV) in the contract shall not be less than 25% of the total value of the contract.
10	Lot 12-1 Waste Management Unit (Microwave based) You have stated specifications that the processing capacity to be minimum 50 250 kg/hour, suitable for high-volume medical waste. There are different capacities from 50kg, 100kg to 250kg,	The minimum required capacity is 150kg/h.

	which both can be suitable for high volume waste. Therefore, kindly clarify the specific processing capacity required to be offered in the bid.	
11	Manufacturer Authorization sent by email in PDF format from the manufacturer are admissible We can submit product catalogues in soft copy (flash disks) because the catalogues are bulky and the manufacturer no longer provides hard copies in compliance with their policy to save the environment by reducing printing material.	Electronic Bidding is not Permitted as per paragraph 22.1 of ITB and The State Department will Not be held liable for completeness and/or accessibility of any soft document submitted whether in form of brochures or otherwise to support the Bid
12	The tender includes some items that are outside our scope of supply. Would it be acceptable for bidders to submit bids for individual items or selected items within a lot, rather than being required to quote for the entire lot	The tender will be evaluated on a Lot-by-Lot basis and only a single bidder can be awarded any of the Lots
13	Would the end user consider solutions offering short-read technology only, provided they meet the majority of the applications and technical requirements outlined in the tender	Refer to technical specifications page 121 in the bid document.
14	Extension of Opening/Closing Time	The Tender submission date has been extended by 1(One) week up to 31st July 2026 at 10.00am.

All other tendering Terms and Conditions remain unchanged. Any inconvenience caused is highly regretted.



Dr. Ouma Oluga, OGW
PRINCIPAL SECRETARY